

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/02/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:41
INTERNAL STANDARD	VERIFIED	15:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:42
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:45

PO #427

