

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/16/2022

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:42
INTERNAL STANDARD	VERIFIED	08:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:42
BLANK TEST	0.00	08:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:45



P.O. # 711