

SUPERVISOR TEST




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 06/15/2022

OPERATOR NAME:  
J T COON  
Permit #: 36043  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10



P.O. [Signature] 341