

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/23/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:00
INTERNAL STANDARD	VERIFIED	09:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:01
BLANK TEST	0.00	09:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:02
BLANK TEST	0.00	09:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:03
BLANK TEST	0.00	09:04

PO Steven Bunio #789

