

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/01/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:59
INTERNAL STANDARD	VERIFIED	10:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:59
BLANK TEST	0.00	11:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:00
BLANK TEST	0.00	11:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:02
BLANK TEST	0.00	11:02

PO Steven T Bunio

