

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/25/2022

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:15
INTERNAL STANDARD	VERIFIED	17:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:16
BLANK TEST	0.00	17:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:17
BLANK TEST	0.00	17:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:18
BLANK TEST	0.00	17:19



P.D. [Signature] #719