

SOLUTION CHANGE




-----  
WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506  
-----

Date: 07/31/2022

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:39
INTERNAL STANDARD	VERIFIED	11:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:40
BLANK TEST	0.00	11:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:41
BLANK TEST	0.00	11:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:42
BLANK TEST	0.00	11:43



*Det. Phil B. [Signature]* #241