

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/22/2022

OPERATOR NAME:  
S T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:17
INTERNAL STANDARD	VERIFIED	09:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:18
BLANK TEST	0.00	09:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:19
BLANK TEST	0.00	09:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:20
BLANK TEST	0.00	09:21

*ST Bunio*  
8.22.22

