

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/29/2022

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:07
INTERNAL STANDARD	VERIFIED	04:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:07
BLANK TEST	0.00	04:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:08
BLANK TEST	0.00	04:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:10
BLANK TEST	0.00	04:10



Michael A Campbell 4757