

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/09/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:01
INTERNAL STANDARD	VERIFIED	17:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:01
BLANK TEST	0.00	17:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:03
BLANK TEST	0.00	17:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:04
BLANK TEST	0.00	17:05

Handwritten signature

09/09/2022

