

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/23/2022

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:36
INTERNAL STANDARD	VERIFIED	07:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:39
BLANK TEST	0.00	07:40

*P.O. Kenneth W Smith # 301*

*9/23/2022*

