

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/03/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:23
INTERNAL STANDARD	VERIFIED	14:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:23
BLANK TEST	0.00	14:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:24
BLANK TEST	0.00	14:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:26
BLANK TEST	0.00	14:26

