

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/24/2022

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:35
INTERNAL STANDARD	VERIFIED	03:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:35
BLANK TEST	0.00	03:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:37
BLANK TEST	0.00	03:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:38
BLANK TEST	0.00	03:39


#820

