

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/01/2022

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:32
INTERNAL STANDARD	VERIFIED	09:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:32
BLANK TEST	0.00	09:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:33
BLANK TEST	0.00	09:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:35
BLANK TEST	0.00	09:36

J.D. Bateholts 4714

