

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/14/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:13
INTERNAL STANDARD	VERIFIED	17:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:14
BLANK TEST	0.00	17:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:15
BLANK TEST	0.00	17:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:16
BLANK TEST	0.00	17:17

