

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/12/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:53
INTERNAL STANDARD	VERIFIED	08:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:54
BLANK TEST	0.00	08:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:55
BLANK TEST	0.00	08:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:56
BLANK TEST	0.00	08:57

PC #427

