

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 01/02/2023

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:39
INTERNAL STANDARD	VERIFIED	01:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:39
BLANK TEST	0.00	01:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:40
BLANK TEST	0.00	01:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:41
BLANK TEST	0.00	01:42



*Michael A Campbell* #757