

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/30/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:17
INTERNAL STANDARD	VERIFIED	12:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:17
BLANK TEST	0.00	12:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:19
BLANK TEST	0.00	12:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:20
BLANK TEST	0.00	12:21

