

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/01/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:06
INTERNAL STANDARD	VERIFIED	10:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:06
BLANK TEST	0.00	10:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:07
BLANK TEST	0.00	10:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:08
BLANK TEST	0.00	10:09

PO #427

