

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 02/15/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:21
INTERNAL STANDARD	VERIFIED	09:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:22
BLANK TEST	0.00	09:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:23
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:24
BLANK TEST	0.00	09:25

*PO JMF #4127*

