

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/08/2023

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:29
INTERNAL STANDARD	VERIFIED	01:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:30
BLANK TEST	0.00	01:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:31
BLANK TEST	0.00	01:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:32
BLANK TEST	0.00	01:33


#820

