

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/12/2023

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:00
INTERNAL STANDARD	VERIFIED	10:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:01
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:02
BLANK TEST	0.00	10:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:03
BLANK TEST	0.00	10:04

PD Steve Bunio

