

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/19/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:47
INTERNAL STANDARD	VERIFIED	13:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:47
BLANK TEST	0.00	13:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:48
BLANK TEST	0.00	13:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:49
BLANK TEST	0.00	13:50

Donnie J Long
844

