

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/26/2023

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:06
INTERNAL STANDARD	VERIFIED	13:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:06
BLANK TEST	0.00	13:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:07
BLANK TEST	0.00	13:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:09
BLANK TEST	0.00	13:09

Signature #789

