

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/01/2023

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:30
INTERNAL STANDARD	VERIFIED	10:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:30
BLANK TEST	0.00	10:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34

