

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/22/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:36
INTERNAL STANDARD	VERIFIED	09:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:37
BLANK TEST	0.00	09:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:38
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:39
BLANK TEST	0.00	09:40

Donnie J Long
844

