

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/29/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:38
INTERNAL STANDARD	VERIFIED	13:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:39
BLANK TEST	0.00	13:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:40
BLANK TEST	0.00	13:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:41
BLANK TEST	0.00	13:42



John D Bateholts 16714