

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 06/22/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:25
INTERNAL STANDARD	VERIFIED	12:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:25
BLANK TEST	0.00	12:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:26
BLANK TEST	0.00	12:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:28
BLANK TEST	0.00	12:28

*PO* *J. Fish*  
#427

