

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/01/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:28
INTERNAL STANDARD	VERIFIED	09:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:29
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32

PO JMF #407

