

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/16/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:48
INTERNAL STANDARD	VERIFIED	09:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:48
BLANK TEST	0.00	09:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:51
BLANK TEST	0.00	09:51



P.O. [Signature]