

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/23/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 11:50 |
| INTERNAL STANDARD       | VERIFIED | 11:50 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 11:50 |
| BLANK TEST              | 0.00     | 11:51 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 11:51 |
| BLANK TEST              | 0.00     | 11:52 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 11:53 |
| BLANK TEST              | 0.00     | 11:53 |

*[Handwritten signature]* *[Handwritten initials]*

