

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/25/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:02
INTERNAL STANDARD	VERIFIED	17:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:05
BLANK TEST	0.00	17:06

John D Bateholts

