

SOLUTION CHANGE



WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/11/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:43
INTERNAL STANDARD	VERIFIED	09:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:44
BLANK TEST	0.00	09:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:45
BLANK TEST	0.00	09:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:46
BLANK TEST	0.00	09:47



John D Bateholts 4714