

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/15/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:51
INTERNAL STANDARD	VERIFIED	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:51
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:52
BLANK TEST	0.00	11:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:53
BLANK TEST	0.00	11:54

