

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/31/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:49
INTERNAL STANDARD	VERIFIED	09:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:50
BLANK TEST	0.00	09:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:51
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:52
BLANK TEST	0.00	09:53

*PO JMF #477*

