

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/22/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:15
INTERNAL STANDARD	VERIFIED	10:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:15
BLANK TEST	0.00	10:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:16
BLANK TEST	0.00	10:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:18
BLANK TEST	0.00	10:18

