

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/01/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:09
INTERNAL STANDARD	VERIFIED	10:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:09
BLANK TEST	0.00	10:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:10
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:12

Don J Long 844

