

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/14/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:13
INTERNAL STANDARD	VERIFIED	12:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:13
BLANK TEST	0.00	12:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:14
BLANK TEST	0.00	12:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:16
BLANK TEST	0.00	12:16

