

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/29/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:19
INTERNAL STANDARD	VERIFIED	17:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:19
BLANK TEST	0.00	17:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:20
BLANK TEST	0.00	17:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:21
BLANK TEST	0.00	17:22

