

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/08/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:09
INTERNAL STANDARD	VERIFIED	09:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:09
BLANK TEST	0.00	09:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:11
BLANK TEST	0.00	09:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:12
BLANK TEST	0.00	09:13

PO [Signature]

