

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/15/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:36
INTERNAL STANDARD	VERIFIED	09:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:37
BLANK TEST	0.00	09:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:38
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:39
BLANK TEST	0.00	09:40

