

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/15/2023

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:41
INTERNAL STANDARD	VERIFIED	11:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:42
BLANK TEST	0.00	11:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:43
BLANK TEST	0.00	11:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:44
BLANK TEST	0.00	11:45

PO Steve EA#789

