

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 12/22/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:04
INTERNAL STANDARD	VERIFIED	08:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:05
BLANK TEST	0.00	08:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:06
BLANK TEST	0.00	08:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:07
BLANK TEST	0.00	08:08

