

SOLUTION CHANGE

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606




Date: 01/02/2023

OPERATOR NAME:
 J RIEHL
 Permit #: 47503
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:06
INTERNAL STANDARD	VERIFIED	01:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:09
BLANK TEST	0.00	01:10

#820



SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/02/2023

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34

#820

