

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 01/23/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:06
INTERNAL STANDARD	VERIFIED	09:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:07
BLANK TEST	0.00	09:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:09
BLANK TEST	0.00	09:10

*PO #427*

