

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 02/01/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:43
INTERNAL STANDARD	VERIFIED	07:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:46
BLANK TEST	0.00	07:46

*PO Fish #427*

