

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 02/22/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:47
INTERNAL STANDARD	VERIFIED	18:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:48
BLANK TEST	0.00	18:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:50
BLANK TEST	0.00	18:51

