

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 03/08/2023

OPERATOR NAME:
 J RIEHL
 Permit #: 47503
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:27
INTERNAL STANDARD	VERIFIED	02:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:31

[Handwritten signature]
 #880

