

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/24/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:35
INTERNAL STANDARD	VERIFIED	07:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:37
BLANK TEST	0.00	07:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:38
BLANK TEST	0.00	07:39

PO Fish #427

