

SUPERVISOR TEST




WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106606

Date: 04/07/2023

OPERATOR NAME:  
 JOHN D BATEHOLTS  
 Permit #: 43007  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:23
INTERNAL STANDARD	VERIFIED	09:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:24
BLANK TEST	0.00	09:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:26
BLANK TEST	0.00	09:27



*J.D. Bateholts* 0719